

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW102122\
 Data File : VW028654.D
 Acq On : 21 Oct 2022 11:52
 Operator : SY/MD
 Sample : N5216-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHAD2

Quant Time: Oct 21 23:00:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102022WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 21 22:58:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	210897	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	137042	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	59519	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.294	65	100326	4.805	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.000%	
7) Chloroethane-d5	1.558	69	79114	4.820	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.400%	
11) 1,1-Dichloroethene-d2	2.095	63	138806	3.825	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.600%	
20) 2-Butanone-d5	3.957	46	201525	51.280	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.560%	
24) Chloroform-d	4.339	84	161302	5.208	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.200%	
26) 1,2-Dichloroethane-d4	5.027	65	84016	5.008	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.200%	
32) Benzene-d6	5.034	84	331612	5.562	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	111.200%	
36) 1,2-Dichloropropane-d6	6.059	67	110339	5.902	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	118.000%	
41) Toluene-d8	7.304	98	201104	4.663	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
43) trans-1,3-Dichloroprop...	7.616	79	18731	4.127	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.600%	
46) 2-Hexanone-d5	8.108	63	91515	48.561	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.120%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	44740	5.066	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.400%	
66) 1,2-Dichlorobenzene-d4	11.615	152	56574	5.288	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.800%	
Target Compounds						
13) Acetone	2.249	43	11597	4.603	ug/L	90
25) Chloroform	4.365	83	87291	2.657	ug/L	99
42) Toluene	7.378	91	10801	0.203	ug/L	96
53) m,p-Xylene	9.133	106	1979	0.098	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102122\
Data File : VV028654.D
Acq On : 21 Oct 2022 11:52
Operator : SY/MD
Sample : N5216-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHAD2

Quant Time: Oct 21 23:00:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102022WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 21 22:58:36 2022
Response via : Initial Calibration

