

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW102122\  
 Data File : VW028671.D  
 Acq On : 21 Oct 2022 19:08  
 Operator : SY/MD  
 Sample : N5216-17  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 BHA79

Quant Time: Oct 21 23:05:31 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR102022WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Oct 21 22:58:36 2022  
 Response via : Initial Calibration

| Compound                      |        | R.T.   | QIon     | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|--------|----------|----------|--------|----------|----------|
| -----                         |        |        |          |          |        |          |          |
| Internal Standards            |        |        |          |          |        |          |          |
| 1) 1,4-Difluorobenzene        |        | 5.616  | 114      | 171066   | 5.000  | ug/L     | 0.02     |
| 28) Chlorobenzene-d5          |        | 8.847  | 117      | 161319   | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    |        | 11.242 | 152      | 71197    | 5.000  | ug/L     | 0.00     |
|                               |        |        |          |          |        |          |          |
| System Monitoring Compounds   |        |        |          |          |        |          |          |
| 4) Vinyl Chloride-d3          |        | 1.304  | 65       | 86706    | 5.120  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 40 - 130 | Recovery | =      | 102.400% |          |
| 7) Chloroethane-d5            |        | 1.564  | 69       | 76131    | 5.719  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 65 - 130 | Recovery | =      | 114.400% |          |
| 11) 1,1-Dichloroethene-d2     |        | 2.104  | 63       | 124491   | 4.230  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 60 - 125 | Recovery | =      | 84.600%  |          |
| 20) 2-Butanone-d5             |        | 3.886  | 46       | 93718    | 29.400 | ug/L     | -0.06    |
| Spiked Amount                 | 50.000 | Range  | 40 - 130 | Recovery | =      | 58.800%  |          |
| 24) Chloroform-d              |        | 4.342  | 84       | 127382   | 5.071  | ug/L     | 0.01     |
| Spiked Amount                 | 5.000  | Range  | 70 - 125 | Recovery | =      | 101.400% |          |
| 26) 1,2-Dichloroethane-d4     |        | 5.030  | 65       | 59976    | 4.408  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 70 - 130 | Recovery | =      | 88.200%  |          |
| 32) Benzene-d6                |        | 5.046  | 84       | 258586   | 3.684  | ug/L     | 0.02     |
| Spiked Amount                 | 5.000  | Range  | 70 - 125 | Recovery | =      | 73.600%  |          |
| 36) 1,2-Dichloropropane-d6    |        | 6.066  | 67       | 83202    | 3.780  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 60 - 140 | Recovery | =      | 75.600%  |          |
| 41) Toluene-d8                |        | 7.310  | 98       | 214557   | 4.227  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 70 - 130 | Recovery | =      | 84.600%  |          |
| 43) trans-1,3-Dichloroprop... |        | 7.622  | 79       | 19706    | 3.688  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 55 - 130 | Recovery | =      | 73.800%  |          |
| 46) 2-Hexanone-d5             |        | 8.088  | 63       | 112712   | 50.808 | ug/L     | -0.02    |
| Spiked Amount                 | 50.000 | Range  | 45 - 130 | Recovery | =      | 101.620% |          |
| 56) 1,1,2,2-Tetrachloroeth... |        | 10.210 | 84       | 50576    | 4.865  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 65 - 120 | Recovery | =      | 97.400%  |          |
| 66) 1,2-Dichlorobenzene-d4    |        | 11.615 | 152      | 64552    | 5.044  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 80 - 120 | Recovery | =      | 100.800% |          |
|                               |        |        |          |          |        |          |          |
| Target Compounds              |        |        |          |          |        |          | Qvalue   |
| 17) Methyl tert-butyl Ether   |        | 2.764  | 73       | 32778    | 1.086  | ug/L     | 96       |
| 25) Chloroform                |        | 4.371  | 83       | 67301    | 2.526  | ug/L     | 96       |
| 51) Chlorobenzene             |        | 8.879  | 112      | 26740    | 0.730  | ug/L     | 95       |
| -----                         |        |        |          |          |        |          |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV102122\  
Data File : VV028671.D  
Acq On : 21 Oct 2022 19:08  
Operator : SY/MD  
Sample : N5216-17  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BHA79

Quant Time: Oct 21 23:05:31 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR102022WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Oct 21 22:58:36 2022  
Response via : Initial Calibration

