

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110122\
 Data File : VW028812.D
 Acq On : 01 Nov 2022 13:39
 Operator : SY/MD
 Sample : N5301-13DL 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBXA8DL

Quant Time: Nov 02 05:12:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 02 05:10:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	406175	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	365782	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	154927	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	123517	32.079	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	64.160%	
7) Chloroethane-d5	1.564	69	120972	36.752	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	73.500%	
11) 1,1-Dichloroethene-d2	2.104	63	203281	27.754	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	55.500%#	
21) 2-Butanone-d5	3.873	46	266804	95.189	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	95.190%	
24) Chloroform-d	4.342	84	246087	42.375	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.760%	
26) 1,2-Dichloroethane-d4	5.027	65	174761	45.502	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.000%	
32) Benzene-d6	5.043	84	498261	43.635	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.260%	
36) 1,2-Dichloropropane-d6	6.066	67	179821	45.870	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	91.740%	
41) Toluene-d8	7.310	98	428390	42.675	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	85.340%	
43) trans-1,3-Dichloroprop...	7.619	79	61349	40.745	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	81.500%	
47) 2-Hexanone-d5	8.085	63	155997	88.054	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	88.050%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	233242	45.119	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	90.240%	
66) 1,2-Dichlorobenzene-d4	11.615	152	140301	45.878	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.760%	
Target Compounds						
17) trans-1,2-Dichloroethene	2.760	96	11988	4.091	ug/L	Qvalue 93
20) cis-1,2-Dichloroethene	3.905	96	147355	45.576	ug/L	# 97
34) Trichloroethene	5.915	95	21746	7.221	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110122\
Data File : VV028812.D
Acq On : 01 Nov 2022 13:39
Operator : SY/MD
Sample : N5301-13DL 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBXA8DL

Quant Time: Nov 02 05:12:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 02 05:10:28 2022
Response via : Initial Calibration

