

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110122\
 Data File : VW028821.D
 Acq On : 01 Nov 2022 17:39
 Operator : SY/MD
 Sample : N5354-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YBMP7

Quant Time: Nov 02 05:14:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 02 05:10:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	399059	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	360828	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	152803	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	161668	42.736	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	85.480%	
7) Chloroethane-d5	1.564	69	144058	44.546	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	89.100%	
11) 1,1-Dichloroethene-d2	2.105	63	242687	33.725	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	67.440%	
21) 2-Butanone-d5	3.873	46	255286	92.704	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	92.700%	
24) Chloroform-d	4.342	84	266096	46.638	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.280%	
26) 1,2-Dichloroethane-d4	5.027	65	186257	49.360	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.720%	
32) Benzene-d6	5.043	84	543326	48.235	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.460%	
36) 1,2-Dichloropropane-d6	6.066	67	191413	49.497	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	99.000%	
41) Toluene-d8	7.310	98	448886	45.330	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.660%	
43) trans-1,3-Dichloroprop...	7.619	79	65617	44.178	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	88.360%	
47) 2-Hexanone-d5	8.085	63	156129	89.339	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	89.340%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	236256	46.330	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	92.660%	
66) 1,2-Dichlorobenzene-d4	11.615	152	146414	48.543	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.080%	
Target Compounds						
					Qvalue	
13) Acetone	2.159	43	204644	76.753	ug/L	100
34) Trichloroethene	5.915	95	15010	5.053	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110122\
Data File : VV028821.D
Acq On : 01 Nov 2022 17:39
Operator : SY/MD
Sample : N5354-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YBMP7

Quant Time: Nov 02 05:14:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 02 05:10:28 2022
Response via : Initial Calibration





#13

Acetone

Concen: 76.753 ug/L

RT: 2.159 min Scan# 34

Delta R.T. -0.007 min

Lab File: VV028821.D

Acq: 01 Nov 2022 17:39

Instrument :

MSVOA_V

ClientSampleId :





YBMP7

Tgt Ion: 43 Resp: 204644

Ion Ratio Lower Upper

43 100

58 29.8 0.0 59.0



#34

Trichloroethene

Concen: 5.053 ug/L

RT: 5.915 min Scan# 1516

Delta R.T. 0.006 min

Lab File: VV028821.D

Acq: 01 Nov 2022 17:39

Tgt Ion: 95 Resp: 15010

Ion Ratio Lower Upper

95 100

97 63.0 46.2 85.8

132 83.9 63.2 117.4

130 92.0 66.2 123.0

