

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110425\
 Data File : VV039308.D
 Acq On : 04 Nov 2025 12:21
 Operator : SY/MD
 Sample : VV1104WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV1104WBL01

Quant Time: Nov 05 00:32:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.597	168	1064589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.529	114	1972523	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1793346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	828938	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.934	65	878179	58.053	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	116.100%
35) Dibromofluoromethane	4.497	113	782028	50.317	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	100.640%
50) Toluene-d8	7.233	98	2570057	48.608	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	97.220%
62) 4-Bromofluorobenzene	9.998	95	975805	51.821	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	103.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110425\
Data File : VV039308.D
Acq On : 04 Nov 2025 12:21
Operator : SY/MD
Sample : VV1104WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV1104WBL01

Quant Time: Nov 05 00:32:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

