

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV110425\  
 Data File : VV039316.D  
 Acq On : 04 Nov 2025 15:54  
 Operator : SY/MD  
 Sample : Q3534-09  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 FB-1

Quant Time: Nov 05 00:38:20 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V100725W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 08 05:01:26 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|-------|-----------|
| -----                       |        |       |          |          |       |           |
| Internal Standards          |        |       |          |          |       |           |
| 1) Pentafluorobenzene       | 4.600  | 168   | 711203   | 50.000   | ug/l  | 0.00      |
| 34) 1,4-Difluorobenzene     | 5.532  | 114   | 1312588  | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 8.773  | 117   | 1267623  | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 11.172 | 152   | 578801   | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds |        |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 4.937  | 65    | 620023   | 61.353   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | =     | 122.700%# |
| 35) Dibromofluoromethane    | 4.500  | 113   | 549147   | 53.097   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | =     | 106.200%  |
| 50) Toluene-d8              | 7.236  | 98    | 1720807  | 48.909   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | =     | 97.820%   |
| 62) 4-Bromofluorobenzene    | 10.001 | 95    | 596656   | 47.617   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | =     | 95.240%   |
| Target Compounds            |        |       |          |          |       |           |
|                             |        |       |          |          |       | Qvalue    |
| 30) Chloroform              | 4.291  | 83    | 14184    | 0.761    | ug/l  | 96        |
| 65) Chlorobenzene           | 8.812  | 112   | 17643    | 0.588    | ug/l  | 95        |
| -----                       |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV110425\  
Data File : VV039316.D  
Acq On : 04 Nov 2025 15:54  
Operator : SY/MD  
Sample : Q3534-09  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
FB-1

Quant Time: Nov 05 00:38:20 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V100725W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 08 05:01:26 2025  
Response via : Initial Calibration

