

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110425\
 Data File : VV039320.D
 Acq On : 04 Nov 2025 17:24
 Operator : SY/MD
 Sample : Q3534-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-EB-3

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 11/05/2025
 Supervised By :Mahesh Dadoda 11/05/2025

Quant Time: Nov 05 00:40:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	890347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1641332	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1514755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	687973	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	687885	54.372	ug/l	0.00
Spiked Amount 50.000	Range 81 - 118		Recovery = 108.740%			
35) Dibromofluoromethane	4.500	113	619545	47.906	ug/l	0.00
Spiked Amount 50.000	Range 80 - 119		Recovery = 95.820%			
50) Toluene-d8	7.236	98	2004989	45.572	ug/l	0.00
Spiked Amount 50.000	Range 89 - 112		Recovery = 91.140%			
62) 4-Bromofluorobenzene	9.998	95	679654	43.377	ug/l	0.00
Spiked Amount 50.000	Range 85 - 114		Recovery = 86.760%			
Target Compounds						Qvalue
8) Diethyl Ether	1.921	74	7972	1.080	ug/l	90
29) Tetrahydrofuran	4.272	42	9654m	1.952	ug/l	
65) Chlorobenzene	8.805	112	209972	5.855	ug/l	97
88) 1,4-Dichlorobenzene	11.194	146	26527	1.035	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110425\
Data File : VV039320.D
Acq On : 04 Nov 2025 17:24
Operator : SY/MD
Sample : Q3534-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-EB-3

Manual Integrations
APPROVED

Reviewed By :Amit Patel 11/05/2025
Supervised By :Mahesh Dadoda 11/05/2025

Quant Time: Nov 05 00:40:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

