

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110425\
 Data File : VV039323.D
 Acq On : 04 Nov 2025 18:32
 Operator : SY/MD
 Sample : Q3534-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-11

Quant Time: Nov 05 00:41:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	635366	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1165878	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1122738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	520406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	575110	63.701	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 127.400%#	
35) Dibromofluoromethane	4.500	113	504189	54.885	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 109.760%	
50) Toluene-d8	7.236	98	1571852	50.297	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 100.600%	
62) 4-Bromofluorobenzene	9.998	95	546785	49.128	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 98.260%	
Target Compounds						
11) Tert butyl alcohol	2.571	59	10599	6.058	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110425\
Data File : VV039323.D
Acq On : 04 Nov 2025 18:32
Operator : SY/MD
Sample : Q3534-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-11

Quant Time: Nov 05 00:41:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

