

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112322\
 Data File : VW029338.D
 Acq On : 24 Nov 2022 09:43
 Operator : SY/MD
 Sample : N5725-09DL 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHB65DL

Quant Time: Nov 25 00:24:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 24 04:08:22 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	212972	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	177923	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	80475	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	74519	5.686	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	113.800%	
7) Chloroethane-d5	1.565	69	64024	6.673	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	133.400%#	
11) 1,1-Dichloroethene-d2	2.105	63	88169	4.204	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	84.000%	
20) 2-Butanone-d5	3.912	46	150234	68.521	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	137.040%#	
24) Chloroform-d	4.339	84	149313	6.383	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	127.600%#	
26) 1,2-Dichloroethane-d4	5.027	65	73115	7.103	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	142.000%#	
32) Benzene-d6	5.044	84	314424	6.394	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	127.800%#	
36) 1,2-Dichloropropane-d6	6.063	67	96773	7.004	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	140.000%	
41) Toluene-d8	7.310	98	258016	5.559	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	111.200%	
43) trans-1,3-Dichloroprop...	7.616	79	32911	6.178	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	123.600%	
46) 2-Hexanone-d5	8.085	63	150411	79.208	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	158.420%#	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	64304	7.542	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	150.800%#	
66) 1,2-Dichlorobenzene-d4	11.616	152	89877	6.393	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	127.800%#	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.114	96	8205	0.749	ug/L #	1
16) Methylene chloride	2.503	84	19848	1.186	ug/L	98
19) 1,1-Dichloroethane	3.185	63	35907	1.587	ug/L	97
22) cis-1,2-Dichloroethene	3.905	96	130057	9.047	ug/L	96
29) 1,1,1-Trichloroethane	4.600	97	35353	1.727	ug/L	100
34) Trichloroethene	5.912	95	21898	1.613	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112322\
Data File : VV029338.D
Acq On : 24 Nov 2022 09:43
Operator : SY/MD
Sample : N5725-09DL 20X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHB65DL

Quant Time: Nov 25 00:24:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 24 04:08:22 2022
Response via : Initial Calibration

