

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112322\
 Data File : VW029339.D
 Acq On : 24 Nov 2022 10:08
 Operator : SY/MD
 Sample : N5725-10DL 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHB66DL

Quant Time: Nov 25 00:24:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 24 04:08:22 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	245870	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	215702	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	111061	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	81949	5.416	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 108.400%		
7) Chloroethane-d5	1.564	69	65519	5.915	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 118.400%		
11) 1,1-Dichloroethene-d2	2.105	63	105133	4.342	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 86.800%		
20) 2-Butanone-d5	3.899	46	125839	49.715	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 99.440%		
24) Chloroform-d	4.339	84	149950	5.552	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 111.000%		
26) 1,2-Dichloroethane-d4	5.027	65	71237	5.995	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 119.800%		
32) Benzene-d6	5.043	84	329541	5.528	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 110.600%		
36) 1,2-Dichloropropane-d6	6.063	67	96986	5.790	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 115.800%		
41) Toluene-d8	7.307	98	308558	5.484	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 109.600%		
43) trans-1,3-Dichloroprop...	7.616	79	32394	5.016	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 100.400%		
46) 2-Hexanone-d5	8.085	63	132350	57.490	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 114.980%		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	61349	5.935	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 118.800%		
66) 1,2-Dichlorobenzene-d4	11.612	152	109578	5.648	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 113.000%		
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.114	96	9777	0.773	ug/L	# 1
16) Methylene chloride	2.503	84	16678	0.863	ug/L	94
19) 1,1-Dichloroethane	3.185	63	36015	1.379	ug/L	97
22) cis-1,2-Dichloroethene	3.902	96	132409	7.978	ug/L	97
29) 1,1,1-Trichloroethane	4.600	97	41788	1.684	ug/L	99
34) Trichloroethene	5.908	95	27164	1.651	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112322\
Data File : VV029339.D
Acq On : 24 Nov 2022 10:08
Operator : SY/MD
Sample : N5725-10DL 20X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHB66DL

Quant Time: Nov 25 00:24:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 24 04:08:22 2022
Response via : Initial Calibration

