

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112322\
 Data File : VV029316.D
 Acq On : 24 Nov 2022 00:20
 Operator : SY/MD
 Sample : N5652-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 24 01:53:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 24 01:40:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	274871	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	246392	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	118735	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	86132	5.092	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.800%
7) Chloroethane-d5	1.565	69	64809	5.234	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.600%
11) 1,1-Dichloroethene-d2	2.105	63	103103	3.809	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.200%
20) 2-Butanone-d5	3.902	46	155504	54.953	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.900%
24) Chloroform-d	4.339	84	151820	5.028	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	5.027	65	71551	5.386	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%
32) Benzene-d6	5.043	84	342469	5.029	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
36) 1,2-Dichloropropane-d6	6.063	67	100481	5.251	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.000%
41) Toluene-d8	7.310	98	314865	4.899	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
43) trans-1,3-Dichloroprop...	7.616	79	35910	4.868	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.400%
46) 2-Hexanone-d5	8.085	63	128857	49.001	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.000%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	61150	5.179	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.600%
66) 1,2-Dichlorobenzene-d4	11.612	152	109019	5.256	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112322\
Data File : VV029316.D
Acq On : 24 Nov 2022 00:20
Operator : SY/MD
Sample : N5652-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 24 01:53:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 24 01:40:30 2022
Response via : Initial Calibration

