

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042622\
 Data File : VW025729.D
 Acq On : 26 Apr 2022 13:33
 Operator : SY/MD
 Sample : N2492-04DL 400X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6T6DL

Quant Time: Apr 27 05:35:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 27 05:32:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	150649	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	152427	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	71032	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	59891	4.276	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.600%	
7) Chloroethane-d5	1.571	69	57421	3.854	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.000%	
11) 1,1-Dichloroethene-d2	2.111	63	93120	3.319	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.400%	
20) 2-Butanone-d5	3.902	46	89103	52.300	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.600%	
24) Chloroform-d	4.352	84	104192	4.923	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.400%	
26) 1,2-Dichloroethane-d4	5.037	65	48577	5.327	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.600%	
32) Benzene-d6	5.053	84	201167	4.561	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.200%	
36) 1,2-Dichloropropane-d6	6.072	67	61598	5.008	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.200%	
41) Toluene-d8	7.317	98	168008	4.105	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.000%	
43) trans-1,3-Dichloroprop...	7.625	79	17664	4.273	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	85.400%	
46) 2-Hexanone-d5	8.088	63	65755	45.319	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.640%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35112	4.519	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	90.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	55641	4.672	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.400%	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.516	84	2852	0.225	ug/L	99
47) Tetrachloroethene	7.976	164	37829	4.042	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042622\
Data File : VV025729.D
Acq On : 26 Apr 2022 13:33
Operator : SY/MD
Sample : N2492-04DL 400X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6T6DL

Quant Time: Apr 27 05:35:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 27 05:32:04 2022
Response via : Initial Calibration

