

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060222\
 Data File : VW025987.D
 Acq On : 02 Jun 2022 16:22
 Operator : SY/MD
 Sample : N3171-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBTZ6

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/06/2022
 Supervised By : Mahesh Dadoda 06/06/2022

Quant Time: Jun 03 00:43:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 03 00:06:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	316099	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	313570	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	142744	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	152706	37.432	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	74.860%		
7) Chloroethane-d5	1.571	69	130941	36.549	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	73.100%		
11) 1,1-Dichloroethene-d2	2.111	63	221344	31.022	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.040%		
21) 2-Butanone-d5	3.902	46	137587	78.687	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	78.690%		
24) Chloroform-d	4.352	84	254680	44.696	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.400%		
26) 1,2-Dichloroethane-d4	5.037	65	157021	45.596	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.200%		
32) Benzene-d6	5.053	84	511809	46.819	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.640%		
36) 1,2-Dichloropropane-d6	6.072	67	163551	48.712	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	97.420%		
41) Toluene-d8	7.317	98	444171	44.100	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.200%		
43) trans-1,3-Dichloroprop...	7.622	79	65020	43.382	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	86.760%		
47) 2-Hexanone-d5	8.092	63	118297	96.103	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	96.100%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	219248	48.967	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	97.940%		
66) 1,2-Dichlorobenzene-d4	11.622	152	163621	52.255	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	104.520%		
Target Compounds					Qvalue	
13) Acetone	2.191	43	11219m	9.402	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060222\
Data File : VV025987.D
Acq On : 02 Jun 2022 16:22
Operator : SY/MD
Sample : N3171-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_V
ClientSampleId :
DBTZ6

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/06/2022
Supervised By : Mahesh Dadoda 06/06/2022


Quant Time: Jun 03 00:43:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 03 00:06:32 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 9.402 ug/L m
 RT: 2.191 min Scan# 3
 Delta R.T. -0.003 min
 Lab File: VV025987.D
 Acq: 02 Jun 2022 16:22

Instrument :
 MSVOA_V
 ClientSampleId :
 DBTZ6



Tgt Ion: 43 Resp: 11219

Ion	Ratio	Lower	Upper
43	100		
58	26.7	0.0	65.0

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/06/2022
 Supervised By :Mahesh Dadoda 06/06/2022

