

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061622\
 Data File : VW026325.D
 Acq On : 16 Jun 2022 11:47
 Operator : SY/MD
 Sample : N3273-09
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B79

Quant Time: Jun 17 05:19:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 05:16:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	191131	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	184001	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73517	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	73563	4.431	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.600%
7) Chloroethane-d5	1.580	69	63281	4.848	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.121	63	103009	3.613	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.200%
20) 2-Butanone-d5	3.921	46	77132	47.616	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.240%
24) Chloroform-d	4.362	84	126549	5.201	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	5.043	65	49624	4.821	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.059	84	243564	4.951	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	6.075	67	73175	5.097	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.000%
41) Toluene-d8	7.320	98	192621	4.294	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
43) trans-1,3-Dichloroprop...	7.628	79	19017	4.085	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.091	63	68991	49.058	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.120%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	40713	4.603	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	62854	5.550	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.000%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.767	101	254084	11.157	ug/L	99
13) Acetone	2.214	43	33373	26.584	ug/L	90
31) Carbon tetrachloride	4.841	117	5387	0.313	ug/L	98
34) Trichloroethene	5.924	95	35857	2.784	ug/L	98
47) Tetrachloroethene	7.976	164	1353213	141.853	ug/L	98
51) Chlorobenzene	8.886	112	2686	0.083	ug/L #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061622\
Data File : VV026325.D
Acq On : 16 Jun 2022 11:47
Operator : SY/MD
Sample : N3273-09
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B79

Quant Time: Jun 17 05:19:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 05:16:46 2022
Response via : Initial Calibration

