

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW061622\  
 Data File : VW026337.D  
 Acq On : 16 Jun 2022 16:33  
 Operator : SY/MD  
 Sample : N3275-09  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 C0BE1

Quant Time: Jun 17 05:22:34 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR061522WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Jun 17 05:16:46 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 5.625  | 114            | 194970   | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 8.853  | 117            | 188763   | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152            | 74598    | 5.000  | ug/L     | 0.00     |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.320  | 65             | 68190    | 4.027  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 80.600%  |          |
| 7) Chloroethane-d5            | 1.580  | 69             | 57905    | 4.349  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 87.000%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.121  | 63             | 94445    | 3.248  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 65.000%  |          |
| 20) 2-Butanone-d5             | 3.928  | 46             | 89534    | 54.184 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 108.360% |          |
| 24) Chloroform-d              | 4.362  | 84             | 124068   | 4.999  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 100.000% |          |
| 26) 1,2-Dichloroethane-d4     | 5.047  | 65             | 49906    | 4.752  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 95.000%  |          |
| 32) Benzene-d6                | 5.059  | 84             | 232608   | 4.609  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 92.200%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.075  | 67             | 71108    | 4.828  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 96.600%  |          |
| 41) Toluene-d8                | 7.320  | 98             | 184587   | 4.011  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 80.200%  |          |
| 43) trans-1,3-Dichloroprop... | 7.628  | 79             | 18006    | 3.770  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 75.400%  |          |
| 46) 2-Hexanone-d5             | 8.091  | 63             | 61818    | 42.848 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 85.700%  |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84             | 42257    | 4.657  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 93.200%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152            | 59203    | 5.152  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 103.000% |          |
| Target Compounds              |        |                |          |        |          |          |
|                               |        |                |          |        |          | Qvalue   |
| 9) Trichlorofluoromethane     | 1.767  | 101            | 3350     | 0.144  | ug/L     | 97       |
| 13) Acetone                   | 2.211  | 43             | 82664    | 64.552 | ug/L     | 85       |
| 34) Trichloroethene           | 5.931  | 95             | 7099     | 0.537  | ug/L     | 98       |
| 47) Tetrachloroethene         | 7.982  | 164            | 1060     | 0.108  | ug/L     | 92       |
| 52) Ethylbenzene              | 9.027  | 91             | 1792     | 0.035  | ug/L     | 97       |
| 53) m,p-Xylene                | 9.146  | 106            | 3206     | 0.161  | ug/L     | 80       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV061622\  
Data File : VV026337.D  
Acq On : 16 Jun 2022 16:33  
Operator : SY/MD  
Sample : N3275-09  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
C0BE1

Quant Time: Jun 17 05:22:34 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR061522WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Jun 17 05:16:46 2022  
Response via : Initial Calibration

