

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061622\
 Data File : VW026340.D
 Acq On : 16 Jun 2022 18:08
 Operator : SY/MD
 Sample : N3275-13
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BE5

Quant Time: Jun 17 05:23:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 05:16:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	186221	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	177418	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	70519	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	65864	4.072	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	81.400%	
7) Chloroethane-d5	1.580	69	57159	4.494	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.800%	
11) 1,1-Dichloroethene-d2	2.121	63	91171	3.282	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	65.600%	
20) 2-Butanone-d5	3.915	46	93308	59.121	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.240%	
24) Chloroform-d	4.362	84	121143	5.110	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.200%	
26) 1,2-Dichloroethane-d4	5.043	65	50550	5.040	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.800%	
32) Benzene-d6	5.059	84	229100	4.830	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.600%	
36) 1,2-Dichloropropane-d6	6.075	67	71421	5.159	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.200%	
41) Toluene-d8	7.320	98	182985	4.231	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.600%	
43) trans-1,3-Dichloroprop...	7.628	79	17921	3.992	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.800%	
46) 2-Hexanone-d5	8.091	63	64686	47.703	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.400%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43071	5.051	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	59221	5.452	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	109.000%	
Target Compounds						
9) Trichlorofluoromethane	1.767	101	12598	0.568	ug/L	97
13) Acetone	2.204	43	188538	154.147	ug/L	87
34) Trichloroethene	5.928	95	9943	0.801	ug/L	99
47) Tetrachloroethene	7.976	164	14873	1.617	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061622\
Data File : VV026340.D
Acq On : 16 Jun 2022 18:08
Operator : SY/MD
Sample : N3275-13
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BE5

Quant Time: Jun 17 05:23:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 05:16:46 2022
Response via : Initial Calibration

