

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061622\
 Data File : VW026344.D
 Acq On : 16 Jun 2022 19:41
 Operator : SY/MD
 Sample : N3320-06 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGJ80

Quant Time: Jun 17 05:24:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 05:16:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	202331	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	190480	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	76690	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	66648	3.792	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.800%
7) Chloroethane-d5	1.581	69	56506	4.089	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.800%
11) 1,1-Dichloroethene-d2	2.117	63	132537	4.392	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.800%
20) 2-Butanone-d5	3.921	46	85171	49.668	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.340%
24) Chloroform-d	4.359	84	122125	4.741	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	5.043	65	49965	4.585	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.600%
32) Benzene-d6	5.059	84	236783	4.650	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	6.075	67	70205	4.724	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.320	98	189152	4.073	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.400%
43) trans-1,3-Dichloroprop...	7.628	79	16748	3.475	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.600%
46) 2-Hexanone-d5	8.095	63	65896	45.263	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.520%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44404	4.850	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	60393	5.113	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.200%
Target Compounds						
3) Chloromethane	1.256	50	1395	0.083	ug/L	93
5) Vinyl chloride	1.323	62	12887	0.738	ug/L	89
12) 1,1-Dichloroethene	2.127	96	63229	4.946	ug/L	95
14) Carbon disulfide	2.304	76	4296	0.136	ug/L	98
15) Methyl Acetate	2.458	43	5572	1.747	ug/L	95
16) Methylene chloride	2.519	84	13819	0.847	ug/L	96
19) 1,1-Dichloroethane	3.201	63	46501	1.967	ug/L	98
22) cis-1,2-Dichloroethene	3.924	96	83824	6.683	ug/L	98
34) Trichloroethene	5.924	95	10829	0.812	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061622\
Data File : VV026344.D
Acq On : 16 Jun 2022 19:41
Operator : SY/MD
Sample : N3320-06 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGJ80

Quant Time: Jun 17 05:24:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 05:16:46 2022
Response via : Initial Calibration

