

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061622\
 Data File : VV026360.D
 Acq On : 17 Jun 2022 02:28
 Operator : SY/MD
 Sample : N3318-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGJ56

Quant Time: Jun 17 07:22:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 06:57:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	186062	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	178285	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73775	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	62309	3.856	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	77.200%	
7) Chloroethane-d5	1.584	69	53406	4.203	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.000%	
11) 1,1-Dichloroethene-d2	2.124	63	105199	3.791	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	75.800%	
20) 2-Butanone-d5	3.928	46	85226	54.046	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.100%	
24) Chloroform-d	4.362	84	117847	4.975	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4	5.043	65	49624	4.952	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.000%	
32) Benzene-d6	5.059	84	224576	4.712	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.200%	
36) 1,2-Dichloropropane-d6	6.076	67	71123	5.113	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.200%	
41) Toluene-d8	7.320	98	178495	4.107	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.200%	
43) trans-1,3-Dichloroprop...	7.629	79	13742	3.046	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	61.000%	
46) 2-Hexanone-d5	8.095	63	61373	45.040	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.080%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44458	5.188	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	58376	5.137	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.800%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.327	62	47223	2.940	ug/L #	81
12) 1,1-Dichloroethene	2.134	96	35607	3.029	ug/L #	80
17) Methyl tert-butyl Ether	2.786	73	1611	0.080	ug/L #	87
18) trans-1,2-Dichloroethene	2.783	96	2738	0.240	ug/L	84
19) 1,1-Dichloroethane	3.208	63	33147	1.525	ug/L	99
22) cis-1,2-Dichloroethene	3.928	96	311777	27.030	ug/L	100
34) Trichloroethene	5.918	95	155340	12.448	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061622\
Data File : VV026360.D
Acq On : 17 Jun 2022 02:28
Operator : SY/MD
Sample : N3318-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGJ56

Quant Time: Jun 17 07:22:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 06:57:04 2022
Response via : Initial Calibration

