

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061622\
 Data File : VV026367.D
 Acq On : 17 Jun 2022 05:15
 Operator : SY/MD
 Sample : N3318-15 10X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGJ69

Quant Time: Jun 17 07:23:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 06:57:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	194643	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	186877	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74254	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	63026	3.728	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.600%
7) Chloroethane-d5	1.584	69	56809	4.274	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.400%
11) 1,1-Dichloroethene-d2	2.124	63	99589	3.430	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.600%
20) 2-Butanone-d5	3.924	46	76217	46.202	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.400%
24) Chloroform-d	4.362	84	122327	4.937	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
26) 1,2-Dichloroethane-d4	5.043	65	48522	4.628	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.059	84	227822	4.560	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
36) 1,2-Dichloropropane-d6	6.075	67	70904	4.863	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.200%
41) Toluene-d8	7.320	98	177149	3.889	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.800%
43) trans-1,3-Dichloroprop...	7.628	79	12526	2.649	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	53.000%#
46) 2-Hexanone-d5	8.091	63	57792	40.462	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.920%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43288	4.819	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	57301	5.010	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.200%
Target Compounds						
5) Vinyl chloride	1.326	62	40922	2.436	ug/L #	79
12) 1,1-Dichloroethene	2.133	96	16126	1.311	ug/L #	7
14) Carbon disulfide	2.310	76	7433	0.245	ug/L	99
15) Methyl Acetate	2.468	43	2260	0.737	ug/L #	62
16) Methylene chloride	2.526	84	6143	0.392	ug/L	88
19) 1,1-Dichloroethane	3.207	63	33742	1.484	ug/L	99
22) cis-1,2-Dichloroethene	3.931	96	50330	4.171	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061622\
Data File : VV026367.D
Acq On : 17 Jun 2022 05:15
Operator : SY/MD
Sample : N3318-15 10X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGJ69

Quant Time: Jun 17 07:23:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 06:57:04 2022
Response via : Initial Calibration

