

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081722\
 Data File : VV027415.D
 Acq On : 17 Aug 2022 10:31
 Operator : SY/MD
 Sample : VV0817WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK206

Quant Time: Aug 18 01:41:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Aug 18 01:40:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	124782	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	138677	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56895	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	39583	3.651	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.000%
7) Chloroethane-d5	1.564	69	35402	4.496	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.000%
11) 1,1-Dichloroethene-d2	2.105	63	67645	3.441	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.800%
20) 2-Butanone-d5	3.889	46	88549	48.229	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.460%
24) Chloroform-d	4.346	84	99662	5.013	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.031	65	53734	5.673	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.400%
32) Benzene-d6	5.047	84	175226	4.283	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	6.066	67	66093	5.225	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.600%
41) Toluene-d8	7.313	98	136645	3.649	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.000%
43) trans-1,3-Dichloroprop...	7.625	79	20909	4.772	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.400%
46) 2-Hexanone-d5	8.088	63	89761	52.939	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.880%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	49670	5.428	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	51987	5.294	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV081722\
Data File : VV027415.D
Acq On : 17 Aug 2022 10:31
Operator : SY/MD
Sample : VV0817WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK206

Quant Time: Aug 18 01:41:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Aug 18 01:40:54 2022
Response via : Initial Calibration

