

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091922\
 Data File : VV027943.D
 Acq On : 19 Sep 2022 22:42
 Operator : SY/MD
 Sample : N4658-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E96

Quant Time: Sep 20 02:11:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Sep 20 01:54:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	112997	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	120162	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	54463	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	44234	5.013	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.200%
7) Chloroethane-d5	1.565	69	39152	5.044	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.800%
11) 1,1-Dichloroethene-d2	2.101	63	54263	3.720	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	3.889	46	76726	54.423	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.840%
24) Chloroform-d	4.343	84	74615	5.051	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
26) 1,2-Dichloroethane-d4	5.031	65	34217	5.598	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.000%
32) Benzene-d6	5.043	84	141292	4.958	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	6.066	67	47415	5.068	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.313	98	110901	4.336	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	7.622	79	13734	4.433	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.600%
46) 2-Hexanone-d5	8.088	63	67394	49.263	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.520%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	37542	4.906	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	45916	5.417	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091922\
Data File : VV027943.D
Acq On : 19 Sep 2022 22:42
Operator : SY/MD
Sample : N4658-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E96

Quant Time: Sep 20 02:11:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Sep 20 01:54:53 2022
Response via : Initial Calibration

