

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091922\
 Data File : VV027936.D
 Acq On : 19 Sep 2022 19:49
 Operator : SY/MD
 Sample : VV0919WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK232

Quant Time: Sep 20 01:55:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Sep 20 01:54:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	118336	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	115961	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	49491	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	47838	5.177	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.600%
7) Chloroethane-d5	1.564	69	40901	5.032	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.600%
11) 1,1-Dichloroethene-d2	2.101	63	59936	3.923	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.400%
20) 2-Butanone-d5	3.889	46	83229	56.372	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.740%
24) Chloroform-d	4.342	84	78662	5.085	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
26) 1,2-Dichloroethane-d4	5.027	65	34756	5.430	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.600%
32) Benzene-d6	5.043	84	147342	5.358	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.200%
36) 1,2-Dichloropropane-d6	6.066	67	49700	5.505	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.200%
41) Toluene-d8	7.313	98	118494	4.801	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
43) trans-1,3-Dichloroprop...	7.622	79	15375	5.142	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.800%
46) 2-Hexanone-d5	8.088	63	71783	54.373	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.740%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	34650	4.692	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	45277	5.878	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091922\
Data File : VV027936.D
Acq On : 19 Sep 2022 19:49
Operator : SY/MD
Sample : VV0919WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK232

Quant Time: Sep 20 01:55:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Sep 20 01:54:53 2022
Response via : Initial Calibration

