

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091922\
 Data File : VW027951.D
 Acq On : 20 Sep 2022 02:00
 Operator : SY/MD
 Sample : N4658-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EE1

Quant Time: Sep 20 02:48:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Sep 20 01:54:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	107109	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	102284	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	47006	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	41953	5.016	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.400%
7) Chloroethane-d5	1.564	69	37533	5.102	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.000%
11) 1,1-Dichloroethene-d2	2.105	63	53713	3.885	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.600%
20) 2-Butanone-d5	3.899	46	77104	57.698	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.400%
24) Chloroform-d	4.346	84	74859	5.346	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
26) 1,2-Dichloroethane-d4	5.030	65	34070	5.881	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.600%
32) Benzene-d6	5.047	84	137505	5.669	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.400%
36) 1,2-Dichloropropane-d6	6.066	67	43546	5.469	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.400%
41) Toluene-d8	7.313	98	97552	4.481	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	7.625	79	12784	4.848	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.000%
46) 2-Hexanone-d5	8.088	63	62259	53.464	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.920%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33796	5.189	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	40186	5.493	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091922\
Data File : VV027951.D
Acq On : 20 Sep 2022 02:00
Operator : SY/MD
Sample : N4658-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EE1

Quant Time: Sep 20 02:48:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Sep 20 01:54:53 2022
Response via : Initial Calibration

