

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
 Data File : VV027968.D
 Acq On : 20 Sep 2022 17:34
 Operator : SY/MD
 Sample : N4619-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E17

Quant Time: Sep 21 05:16:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 21 01:52:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	355085	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	340505	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	149972	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	145717	46.745	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.500%
7) Chloroethane-d5	1.558	69	114928	46.152	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.300%
11) 1,1-Dichloroethene-d2	2.098	63	195381	35.380	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.760%
21) 2-Butanone-d5	3.873	46	195071	105.236	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.240%
24) Chloroform-d	4.339	84	255136	48.749	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.500%
26) 1,2-Dichloroethane-d4	5.024	65	169496	49.974	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.940%
32) Benzene-d6	5.040	84	498216	50.764	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.520%
36) 1,2-Dichloropropane-d6	6.063	67	163022	53.792	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.580%
41) Toluene-d8	7.310	98	404999	45.063	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.120%
43) trans-1,3-Dichloroprop...	7.619	79	67402	44.475	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.960%
47) 2-Hexanone-d5	8.085	63	147246	99.178	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.180%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	226188	50.707	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.420%
66) 1,2-Dichlorobenzene-d4	11.622	152	153777	52.821	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.640%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.912	95	20213	8.029	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
Data File : VV027968.D
Acq On : 20 Sep 2022 17:34
Operator : SY/MD
Sample : N4619-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E17

Quant Time: Sep 21 05:16:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 21 01:52:39 2022
Response via : Initial Calibration

