

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092022\
 Data File : VW027988.D
 Acq On : 21 Sep 2022 02:15
 Operator : SY/MD
 Sample : N4674-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E95

Quant Time: Sep 21 08:04:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 21 06:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	354831	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	343075	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	148955	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	140728	45.177	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	90.360%		
7) Chloroethane-d5	1.561	69	117227	47.109	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	94.220%		
11) 1,1-Dichloroethene-d2	2.101	63	192762	34.931	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	69.860%		
21) 2-Butanone-d5	3.876	46	212080	114.494	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	114.490%		
24) Chloroform-d	4.339	84	259770	49.670	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.340%		
26) 1,2-Dichloroethane-d4	5.024	65	173318	51.137	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.280%		
32) Benzene-d6	5.043	84	502397	50.807	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.620%		
36) 1,2-Dichloropropane-d6	6.063	67	165569	54.223	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	108.440%		
41) Toluene-d8	7.313	98	423762	46.797	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.600%		
43) trans-1,3-Dichloroprop...	7.619	79	69068	45.233	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	90.460%		
47) 2-Hexanone-d5	8.085	63	152887	102.206	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	102.210%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	232173	51.659	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	103.320%		
66) 1,2-Dichlorobenzene-d4	11.622	152	157101	54.331	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	108.660%		
Target Compounds					Qvalue	
34) Trichloroethene	5.915	95	12547	4.947	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
Data File : VV027988.D
Acq On : 21 Sep 2022 02:15
Operator : SY/MD
Sample : N4674-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E95

Quant Time: Sep 21 08:04:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 21 06:45:13 2022
Response via : Initial Calibration

