

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
 Data File : VV028002.D
 Acq On : 21 Sep 2022 08:00
 Operator : SY/MD
 Sample : N4674-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EB6

Quant Time: Sep 21 08:46:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 21 06:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	336086	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	335260	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	148525	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	120531	40.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.700%
7) Chloroethane-d5	1.561	69	103172	43.774	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.540%
11) 1,1-Dichloroethene-d2	2.101	63	175267	33.532	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.060%
21) 2-Butanone-d5	3.886	46	205950	117.386	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.390%
24) Chloroform-d	4.339	84	246335	49.729	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.460%
26) 1,2-Dichloroethane-d4	5.024	65	165289	51.489	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.980%
32) Benzene-d6	5.043	84	470969	48.739	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.480%
36) 1,2-Dichloropropane-d6	6.063	67	156805	52.550	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.100%
41) Toluene-d8	7.310	98	395999	44.751	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.500%
43) trans-1,3-Dichloroprop...	7.619	79	61576	41.267	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.540%
47) 2-Hexanone-d5	8.085	63	150181	102.738	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.740%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	206313	46.975	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.960%
66) 1,2-Dichlorobenzene-d4	11.622	152	155285	53.859	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.720%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.111	96	13522	5.545	ug/L	# 1
20) cis-1,2-Dichloroethene	3.905	96	12926	4.994	ug/L	82
30) 1,1,1-Trichloroethane	4.606	97	6403	1.551	ug/L	98
34) Trichloroethene	5.908	95	176456	71.192	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
Data File : VV028002.D
Acq On : 21 Sep 2022 08:00
Operator : SY/MD
Sample : N4674-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EB6

Quant Time: Sep 21 08:46:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 21 06:45:13 2022
Response via : Initial Calibration

