

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092022\
 Data File : VW027980.D
 Acq On : 20 Sep 2022 22:32
 Operator : SY/MD
 Sample : N4619-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E35

Quant Time: Sep 21 05:18:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 21 01:52:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	357756	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	342321	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	151998	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	142589	45.401	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.800%
7) Chloroethane-d5	1.558	69	118346	47.170	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.340%
11) 1,1-Dichloroethene-d2	2.098	63	196162	35.256	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.520%
21) 2-Butanone-d5	3.873	46	196177	105.043	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.040%
24) Chloroform-d	4.339	84	242375	45.965	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.940%
26) 1,2-Dichloroethane-d4	5.024	65	166213	48.640	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.280%
32) Benzene-d6	5.043	84	477522	48.397	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.063	67	165450	54.303	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.600%
41) Toluene-d8	7.310	98	420184	46.504	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.000%
43) trans-1,3-Dichloroprop...	7.619	79	73293	48.106	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.220%
47) 2-Hexanone-d5	8.085	63	163715	109.686	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.690%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	233062	51.971	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.940%
66) 1,2-Dichlorobenzene-d4	11.622	152	158745	53.801	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
Data File : VV027980.D
Acq On : 20 Sep 2022 22:32
Operator : SY/MD
Sample : N4619-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E35

Quant Time: Sep 21 05:18:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 21 01:52:39 2022
Response via : Initial Calibration

