

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
 Data File : VV027987.D
 Acq On : 21 Sep 2022 01:50
 Operator : SY/MD
 Sample : VV0920WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK265

Quant Time: Sep 21 06:45:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 21 06:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	355679	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	340580	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	153491	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	147431	47.216	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.440%
7) Chloroethane-d5	1.558	69	123362	49.457	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.920%
11) 1,1-Dichloroethene-d2	2.101	63	207766	37.560	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.120%
21) 2-Butanone-d5	3.873	46	189058	101.822	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.820%
24) Chloroform-d	4.342	84	246663	47.052	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.100%
26) 1,2-Dichloroethane-d4	5.024	65	159693	47.005	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.020%
32) Benzene-d6	5.043	84	481306	49.030	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.060%
36) 1,2-Dichloropropane-d6	6.063	67	147708	48.728	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.460%
41) Toluene-d8	7.310	98	417002	46.388	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.780%
43) trans-1,3-Dichloroprop...	7.619	79	65485	43.201	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.400%
47) 2-Hexanone-d5	8.085	63	150046	101.042	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.040%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	231459	51.877	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.760%
66) 1,2-Dichlorobenzene-d4	11.622	152	160018	53.705	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092022\
Data File : VV027987.D
Acq On : 21 Sep 2022 01:50
Operator : SY/MD
Sample : VV0920WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK265

Quant Time: Sep 21 06:45:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 21 06:45:13 2022
Response via : Initial Calibration

