

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092322\
 Data File : VV028106.D
 Acq On : 23 Sep 2022 16:15
 Operator : SY/MD
 Sample : VIBLK293
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK293

Quant Time: Sep 24 01:28:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 24 01:23:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	171899	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	174138	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	77632	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	75444	4.233	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	84.600%	
7) Chloroethane-d5	1.561	69	65459	4.620	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.400%	
11) 1,1-Dichloroethene-d2	2.101	63	100266	3.244	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.800%	
20) 2-Butanone-d5	3.880	46	163645	54.391	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.780%	
24) Chloroform-d	4.343	84	140796	4.951	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.000%	
26) 1,2-Dichloroethane-d4	5.027	65	69006	5.100	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.000%	
32) Benzene-d6	5.043	84	259118	4.538	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.800%	
36) 1,2-Dichloropropane-d6	6.066	67	87701	4.927	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.600%	
41) Toluene-d8	7.313	98	199829	3.866	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.400%	
43) trans-1,3-Dichloroprop...	7.622	79	27511	4.279	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	85.600%	
46) 2-Hexanone-d5	8.085	63	125507	46.768	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.540%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	59887	4.922	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	77946	5.119	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.400%	
Target Compounds						Qvalue
16) Methylene chloride	2.500	84	5897	0.445	ug/L	98
34) Trichloroethene	5.912	95	31461	2.614	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092322\
Data File : VV028106.D
Acq On : 23 Sep 2022 16:15
Operator : SY/MD
Sample : VIBLK293
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK293

Quant Time: Sep 24 01:28:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 24 01:23:44 2022
Response via : Initial Calibration

