

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092322\
 Data File : VV028091.D
 Acq On : 23 Sep 2022 10:07
 Operator : SY/MD
 Sample : VV0923WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK234

Quant Time: Sep 24 01:24:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 24 01:23:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	182935	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	186541	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	85010	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	90512	4.772	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.400%	
7) Chloroethane-d5	1.561	69	77060	5.111	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.200%	
11) 1,1-Dichloroethene-d2	2.098	63	118947	3.616	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	72.400%	
20) 2-Butanone-d5	3.876	46	156452	48.863	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	97.720%	
24) Chloroform-d	4.336	84	147636	4.878	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.600%	
26) 1,2-Dichloroethane-d4	5.024	65	70998	4.930	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.600%	
32) Benzene-d6	5.040	84	277706	4.540	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.800%	
36) 1,2-Dichloropropane-d6	6.063	67	90084	4.724	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.400%	
41) Toluene-d8	7.310	98	226110	4.084	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.600%	
43) trans-1,3-Dichloroprop...	7.619	79	29891	4.340	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.800%	
46) 2-Hexanone-d5	8.082	63	129829	45.161	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.320%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	63360	4.861	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	84629	5.075	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092322\
Data File : VV028091.D
Acq On : 23 Sep 2022 10:07
Operator : SY/MD
Sample : VV0923WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK234

Quant Time: Sep 24 01:24:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 24 01:23:44 2022
Response via : Initial Calibration

