

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW093022\
 Data File : VW028244.D
 Acq On : 30 Sep 2022 22:53
 Operator : SY/MD
 Sample : N4865-06DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW845DL

Quant Time: Oct 01 02:18:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 29 01:25:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	166456	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	164120	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	75055	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	63552	3.683	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	73.600%	
7) Chloroethane-d5	1.561	69	57304	4.177	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	83.600%	
11) 1,1-Dichloroethene-d2	2.098	63	92966	3.106	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.200%	
20) 2-Butanone-d5	3.905	46	118365	40.628	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	81.260%	
24) Chloroform-d	4.339	84	127193	4.619	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.400%	
26) 1,2-Dichloroethane-d4	5.027	65	58912	4.496	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.000%	
32) Benzene-d6	5.043	84	245428	4.561	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.200%	
36) 1,2-Dichloropropane-d6	6.063	67	75038	4.473	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.400%	
41) Toluene-d8	7.310	98	208547	4.281	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.600%	
43) trans-1,3-Dichloroprop...	7.622	79	23614	3.897	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.000%	
46) 2-Hexanone-d5	8.088	63	113884	45.027	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.060%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	53444	4.661	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	77242	5.247	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.000%	
Target Compounds						
16) Methylene chloride	2.497	84	4146	0.323	ug/L	81
18) trans-1,2-Dichloroethene	2.754	96	2630	0.261	ug/L	84
22) cis-1,2-Dichloroethene	3.905	96	2431	0.184	ug/L	86
34) Trichloroethene	5.918	95	2812	0.248	ug/L	85
47) Tetrachloroethene	7.973	164	86454	9.547	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093022\
Data File : VV028244.D
Acq On : 30 Sep 2022 22:53
Operator : SY/MD
Sample : N4865-06DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW845DL

Quant Time: Oct 01 02:18:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 29 01:25:35 2022
Response via : Initial Calibration

