

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093022\
 Data File : VV028246.D
 Acq On : 30 Sep 2022 23:43
 Operator : SY/MD
 Sample : N4866-02DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXFP8DL

Quant Time: Oct 01 02:18:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 29 01:25:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	165154	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	159702	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	71852	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	62439	3.647	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.000%
7) Chloroethane-d5	1.561	69	54689	4.017	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.400%
11) 1,1-Dichloroethene-d2	2.101	63	90873	3.060	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.200%
20) 2-Butanone-d5	3.896	46	120202	41.584	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.160%
24) Chloroform-d	4.339	84	124492	4.556	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
26) 1,2-Dichloroethane-d4	5.027	65	58261	4.481	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	5.043	84	239392	4.572	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.066	67	74867	4.586	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.800%
41) Toluene-d8	7.313	98	192219	4.055	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
43) trans-1,3-Dichloroprop...	7.622	79	21930	3.719	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.400%
46) 2-Hexanone-d5	8.085	63	109195	44.367	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.740%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	52604	4.714	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	70079	4.972	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.304	62	85961	6.712	ug/L	99
16) Methylene chloride	2.503	84	2710	0.213	ug/L	87
18) trans-1,2-Dichloroethene	2.764	96	1550	0.155	ug/L	71
22) cis-1,2-Dichloroethene	3.905	96	16002	1.221	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093022\
Data File : VV028246.D
Acq On : 30 Sep 2022 23:43
Operator : SY/MD
Sample : N4866-02DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXFP8DL

Quant Time: Oct 01 02:18:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 29 01:25:35 2022
Response via : Initial Calibration

