

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW093022\
 Data File : VW028253.D
 Acq On : 01 Oct 2022 02:38
 Operator : SY/MD
 Sample : N4871-19
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5H82

Quant Time: Oct 01 04:28:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 29 01:25:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	172965	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	175566	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	82855	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	63290	3.529	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.600%
7) Chloroethane-d5	1.561	69	56417	3.957	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.200%
11) 1,1-Dichloroethene-d2	2.098	63	92534	2.975	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.600%#
20) 2-Butanone-d5	3.895	46	145837	48.173	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.340%
24) Chloroform-d	4.339	84	136251	4.762	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
26) 1,2-Dichloroethane-d4	5.027	65	62581	4.596	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.000%
32) Benzene-d6	5.043	84	266513	4.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	6.066	67	84664	4.718	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.313	98	220814	4.238	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.800%
43) trans-1,3-Dichloroprop...	7.622	79	25279	3.899	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.000%
46) 2-Hexanone-d5	8.088	63	119603	44.205	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.420%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	60389	4.923	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	84463	5.197	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093022\
Data File : VV028253.D
Acq On : 01 Oct 2022 02:38
Operator : SY/MD
Sample : N4871-19
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5H82

Quant Time: Oct 01 04:28:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 29 01:25:35 2022
Response via : Initial Calibration

