

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100422\
 Data File : VW028352.D
 Acq On : 05 Oct 2022 02:11
 Operator : SY/MD
 Sample : N4872-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXFX7

Quant Time: Oct 05 04:18:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 04 02:28:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	173521	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	174603	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	83725	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	54472	4.178	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.600%	
7) Chloroethane-d5	1.561	69	52359	4.744	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.800%	
11) 1,1-Dichloroethene-d2	2.101	63	80251	3.266	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	65.400%	
20) 2-Butanone-d5	3.902	46	148882	67.490	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	134.980%#	
24) Chloroform-d	4.342	84	114336	4.841	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
26) 1,2-Dichloroethane-d4	5.027	65	57858	5.284	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.600%	
32) Benzene-d6	5.043	84	240230	5.016	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.400%	
36) 1,2-Dichloropropane-d6	6.066	67	75533	5.497	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	110.000%	
41) Toluene-d8	7.310	98	194875	4.424	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.400%	
43) trans-1,3-Dichloroprop...	7.619	79	23052	4.405	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.200%	
46) 2-Hexanone-d5	8.085	63	116413	54.349	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	108.700%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	55519	5.331	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.600%	
66) 1,2-Dichlorobenzene-d4	11.619	152	82240	5.708	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	114.200%	
Target Compounds						
					Qvalue	
25) Chloroform	4.371	83	9577	0.396	ug/L	98
31) Carbon tetrachloride	4.828	117	3720	0.193	ug/L	98
51) Chlorobenzene	8.879	112	25841	0.710	ug/L	98
64) 1,3-Dichlorobenzene	11.178	146	71933	2.858	ug/L	99
65) 1,4-Dichlorobenzene	11.268	146	206715	8.203	ug/L	99
67) 1,2-Dichlorobenzene	11.638	146	217342	9.522	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100422\
Data File : VV028352.D
Acq On : 05 Oct 2022 02:11
Operator : SY/MD
Sample : N4872-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXFX7

Quant Time: Oct 05 04:18:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 04 02:28:45 2022
Response via : Initial Calibration

