

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011022\
 Data File : VW024319.D
 Acq On : 10 Jan 2022 12:01
 Operator : SY/MD
 Sample : VW0110WBL06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK294

Quant Time: Jan 11 01:52:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR010422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 11 01:51:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	99812	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	99181	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53005	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	33719	4.159	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.200%	
7) Chloroethane-d5	1.568	69	29479	4.626	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.600%	
11) 1,1-Dichloroethene-d2	2.108	63	50235	3.160	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	63.200%	
20) 2-Butanone-d5	3.911	46	47460m	45.038	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	90.080%	
24) Chloroform-d	4.342	84	71484	4.707	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.200%	
26) 1,2-Dichloroethane-d4	5.027	65	35281	4.911	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.200%	
32) Benzene-d6	5.046	84	132631	4.774	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.400%	
36) 1,2-Dichloropropane-d6	6.066	67	37400	5.179	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.600%	
41) Toluene-d8	7.313	98	118460	4.293	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.800%	
43) trans-1,3-Dichloroprop...	7.622	79	14746	4.458	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.200%	
46) 2-Hexanone-d5	8.088	63	49055	47.459	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.920%	
56) 1,1,2,2-Tetrachloroeth...	10.213	84	25983	4.944	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	48896	5.311	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011022\
Data File : VV024319.D
Acq On : 10 Jan 2022 12:01
Operator : SY/MD
Sample : VV0110WBL06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK294

Quant Time: Jan 11 01:52:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR010422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 11 01:51:56 2022
Response via : Initial Calibration

