

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011222\
 Data File : VW024334.D
 Acq On : 12 Jan 2022 16:48
 Operator : SY/MD
 Sample : N1084-03ME
 Misc : 6.44g/5.0mL/100ul/5.0mL/MSVOA_V/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGLM6ME

Quant Time: Jan 13 02:33:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 13 02:28:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	419648	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	389640	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	203704	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	68219	22.588	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	45.180%#
7) Chloroethane-d5	1.568	69	61888	26.975	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	53.940%#
11) 1,1-Dichloroethene-d2	2.108	63	103062	18.259	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	36.520%#
21) 2-Butanone-d5	3.879	46	184895	109.883	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.880%
24) Chloroform-d	4.349	84	208438	37.762	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.520%
26) 1,2-Dichloroethane-d4	5.031	65	139913	37.859	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.720%
32) Benzene-d6	5.050	84	331398	32.593	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	65.180%#
36) 1,2-Dichloropropane-d6	6.069	67	114253	40.094	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.180%
41) Toluene-d8	7.317	98	283930	28.841	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	57.680%#
43) trans-1,3-Dichloroprop...	7.619	79	58918	35.906	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.820%
47) 2-Hexanone-d5	8.085	63	129570	104.658	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.660%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	195583	48.028	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.060%
66) 1,2-Dichlorobenzene-d4	11.622	152	174975	43.094	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011222\
Data File : VV024334.D
Acq On : 12 Jan 2022 16:48
Operator : SY/MD
Sample : N1084-03ME
Misc : 6.44g/5.0mL/100ul/5.0ml/MSVOA_V/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGLM6ME

Quant Time: Jan 13 02:33:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 13 02:28:22 2022
Response via : Initial Calibration

