

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011321\
 Data File : VW019981.D
 Acq On : 13 Jan 2021 12:11
 Operator : SY/MD
 Sample : M1099-04ME 10X
 Misc : 5.60g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F1C04ME

Quant Time: Jan 13 13:07:01 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 13 13:06:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.617	114	404454	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.855	117	385989	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.254	152	201553	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.306	65	171700	50.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.66%
7) Chloroethane-d5	1.566	69	152313	55.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.86%
11) 1,1-Dichloroethene-d2	2.106	63	255747	38.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.12%
21) 2-Butanone-d5	3.881	46	198721	89.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.79%
24) Chloroform-d	4.347	84	282308	45.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.88%
26) 1,2-Dichloroethane-d4	5.032	65	203050	48.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.20%
32) Benzene-d6	5.051	84	523966	48.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.071	67	163532	48.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.16%
41) Toluene-d8	7.318	98	468221	47.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.84%
43) trans-1,3-Dichloroprop...	7.624	79	78722	44.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.20%
47) 2-Hexanone-d5	8.090	63	119397	87.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.56%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	199957	43.49	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.98%
64) 1,2-Dichlorobenzene-d4	11.630	152	187500	48.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.32%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011321\
Data File : VV019981.D
Acq On : 13 Jan 2021 12:11
Operator : SY/MD
Sample : M1099-04ME 10X
Misc : 5.60g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F1C04ME

Quant Time: Jan 13 13:07:01 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM011121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 13 13:06:01 2021
Response via : Initial Calibration

