

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011322\
 Data File : VW024345.D
 Acq On : 13 Jan 2022 11:11
 Operator : SY/MD
 Sample : VW0113MBL01
 Misc : 5.00g/5.0ml/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK292

Quant Time: Jan 13 14:29:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 13 02:28:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	236514	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	222177	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	111701	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	82403	48.411	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.820%
7) Chloroethane-d5	1.568	69	66787	51.650	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.300%
11) 1,1-Dichloroethene-d2	2.108	63	124697	39.197	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.400%
21) 2-Butanone-d5	3.883	46	92242	97.266	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.270%
24) Chloroform-d	4.346	84	160867	51.710	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.420%
26) 1,2-Dichloroethane-d4	5.027	65	107218	51.476	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.960%
32) Benzene-d6	5.047	84	311568	53.739	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.480%
36) 1,2-Dichloropropane-d6	6.066	67	90202	55.513	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	111.020%
41) Toluene-d8	7.313	98	291042	51.846	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.700%
43) trans-1,3-Dichloroprop...	7.619	79	45897	49.053	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.100%
47) 2-Hexanone-d5	8.085	63	65762	93.155	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.150%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	113793	49.005	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	126836	56.967	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011322\
Data File : VV024345.D
Acq On : 13 Jan 2022 11:11
Operator : SY/MD
Sample : VV0113MBL01
Misc : 5.00g/5.0ml/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK292

Quant Time: Jan 13 14:29:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 13 02:28:22 2022
Response via : Initial Calibration

