

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011422\
 Data File : VW024361.D
 Acq On : 14 Jan 2022 15:10
 Operator : SY/MD
 Sample : VW0114WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK294

Quant Time: Jan 17 08:20:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 13 02:28:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	269796	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	254833	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	128020	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	69951	36.026	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.060%
7) Chloroethane-d5	1.568	69	58190	39.450	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.900%
11) 1,1-Dichloroethene-d2	2.108	63	110861	30.549	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.100%
21) 2-Butanone-d5	3.880	46	102350	94.611	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.610%
24) Chloroform-d	4.343	84	153041	43.126	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.260%
26) 1,2-Dichloroethane-d4	5.024	65	101393	42.674	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.340%
32) Benzene-d6	5.043	84	294370	44.266	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.540%
36) 1,2-Dichloropropane-d6	6.063	67	85292	45.765	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.520%
41) Toluene-d8	7.310	98	277724	43.133	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.260%
43) trans-1,3-Dichloroprop...	7.619	79	45235	42.151	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.300%
47) 2-Hexanone-d5	8.085	63	66332	81.921	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.920%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	112039	42.066	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.140%
66) 1,2-Dichlorobenzene-d4	11.622	152	121912	47.776	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011422\
Data File : VV024361.D
Acq On : 14 Jan 2022 15:10
Operator : SY/MD
Sample : VV0114WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK294

Quant Time: Jan 17 08:20:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 13 02:28:22 2022
Response via : Initial Calibration

