

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011722\
 Data File : VW024373.D
 Acq On : 17 Jan 2022 12:28
 Operator : SY/MD
 Sample : N1168-12
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGLX7

Quant Time: Jan 17 14:54:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 13 02:28:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	286500	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	266223	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	130383	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	108406	52.576	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.160%
7) Chloroethane-d5	1.571	69	81656	52.132	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.260%
11) 1,1-Dichloroethene-d2	2.111	63	157020	40.746	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.500%
21) 2-Butanone-d5	3.886	46	142499	124.044	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	124.040%
24) Chloroform-d	4.349	84	217954	57.837	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.680%
26) 1,2-Dichloroethane-d4	5.031	65	140352	55.627	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.260%
32) Benzene-d6	5.050	84	433979	62.468	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	124.940%
36) 1,2-Dichloropropane-d6	6.069	67	128925	66.217	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	132.440%#
41) Toluene-d8	7.314	98	404979	60.207	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	120.420%#
43) trans-1,3-Dichloroprop...	7.622	79	68038	60.686	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	121.380%
47) 2-Hexanone-d5	8.088	63	96472	114.047	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.050%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	161104	57.901	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	115.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	165221	63.574	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	127.140%#
Target Compounds					Qvalue	
13) Acetone	2.179	43	8407	7.412	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011722\
Data File : VV024373.D
Acq On : 17 Jan 2022 12:28
Operator : SY/MD
Sample : N1168-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGLX7

Quant Time: Jan 17 14:54:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 13 02:28:22 2022
Response via : Initial Calibration

