

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011722\
 Data File : VW024371.D
 Acq On : 17 Jan 2022 11:37
 Operator : SY/MD
 Sample : VW0117MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK299

Quant Time: Jan 17 14:54:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 13 02:28:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	309565	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	285691	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	137950	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	105891	47.529	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.060%
7) Chloroethane-d5	1.568	69	78632	46.461	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.920%
11) 1,1-Dichloroethene-d2	2.108	63	150559	36.159	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.320%
21) 2-Butanone-d5	3.876	46	117276	94.481	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.480%
24) Chloroform-d	4.342	84	203586	49.999	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.000%
26) 1,2-Dichloroethane-d4	5.024	65	133328	48.906	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.820%
32) Benzene-d6	5.047	84	411808	55.237	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.480%
36) 1,2-Dichloropropane-d6	6.063	67	124199	59.443	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	118.880%
41) Toluene-d8	7.310	98	384679	53.292	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.580%
43) trans-1,3-Dichloroprop...	7.616	79	62663	52.083	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.160%
47) 2-Hexanone-d5	8.085	63	82099	90.442	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.440%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	151072	50.595	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	154312	56.120	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011722\
Data File : VV024371.D
Acq On : 17 Jan 2022 11:37
Operator : SY/MD
Sample : VV0117MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK299

Quant Time: Jan 17 14:54:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 13 02:28:22 2022
Response via : Initial Calibration

