

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011822\
 Data File : VW024388.D
 Acq On : 18 Jan 2022 12:59
 Operator : SY/MD
 Sample : VIBLK214
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK214

Quant Time: Jan 19 01:04:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 19 01:03:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	244866	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	230702	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	111345	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	87853	49.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.700%
7) Chloroethane-d5	1.571	69	74079	55.336	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	110.680%
11) 1,1-Dichloroethene-d2	2.111	63	133467	40.523	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.040%
21) 2-Butanone-d5	3.889	46	148767	151.519	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	151.520%#
24) Chloroform-d	4.349	84	205126	63.687	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	127.380%#
26) 1,2-Dichloroethane-d4	5.031	65	137765	63.886	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	127.780%#
32) Benzene-d6	5.050	84	404054	67.115	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	134.240%#
36) 1,2-Dichloropropane-d6	6.069	67	126682	75.083	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	150.160%#
41) Toluene-d8	7.313	98	375086	64.348	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	128.700%#
43) trans-1,3-Dichloroprop...	7.622	79	62387	64.214	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	128.420%#
47) 2-Hexanone-d5	8.088	63	99810	136.161	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	136.160%#
56) 1,1,2,2-Tetrachloroeth...	10.214	84	161684	67.056	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	134.120%#
66) 1,2-Dichlorobenzene-d4	11.625	152	156562	70.543	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	141.080%#
Target Compounds						
					Qvalue	
13) Acetone	2.188	43	5201	5.365	ug/L	73
48) 2-Hexanone	8.146	43	11762	8.139	ug/L	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011822\
Data File : VV024388.D
Acq On : 18 Jan 2022 12:59
Operator : SY/MD
Sample : VIBLK214
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK214

Quant Time: Jan 19 01:04:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 19 01:03:04 2022
Response via : Initial Calibration

