

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012424\
 Data File : VW033843.D
 Acq On : 24 Jan 2024 11:10
 Operator : SY/MD
 Sample : P1191-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMQ0

Quant Time: Jan 25 00:56:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 25 00:55:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	212405	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	200555	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	100647	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	83034	50.865	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.740%
7) Chloroethane-d5	1.529	69	65081	50.342	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.680%
11) 1,1-Dichloroethene-d2	2.059	65	35647	50.740	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.480%
21) 2-Butanone-d5	3.783	46	81229	72.545	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	72.550%
24) Chloroform-d	4.249	84	133867	45.861	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.720%
26) 1,2-Dichloroethane-d4	4.940	65	91728	48.636	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.280%
32) Benzene-d6	4.960	84	264054	46.312	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.620%
36) 1,2-Dichloropropane-d6	5.989	67	82182	43.789	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.580%
41) Toluene-d8	7.243	98	245431	47.775	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.560%
43) trans-1,3-Dichloroprop...	7.554	79	41201	44.419	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.840%
47) 2-Hexanone-d5	8.021	63	59159	75.992	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.990%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	101951	42.551	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.100%
66) 1,2-Dichlorobenzene-d4	11.561	152	94155	49.307	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012424\
Data File : VV033843.D
Acq On : 24 Jan 2024 11:10
Operator : SY/MD
Sample : P1191-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMQ0

Quant Time: Jan 25 00:56:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 25 00:55:39 2024
Response via : Initial Calibration

