

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012424\
 Data File : VW033846.D
 Acq On : 24 Jan 2024 12:21
 Operator : SY/MD
 Sample : P1191-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMQ3

Quant Time: Jan 25 00:57:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 25 00:55:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	228397	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	216718	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	111016	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	86481	49.267	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	98.540%		
7) Chloroethane-d5	1.532	69	67545	48.590	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	97.180%		
11) 1,1-Dichloroethene-d2	2.063	65	37666	49.859	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	99.720%		
21) 2-Butanone-d5	3.783	46	84807	70.438	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	70.440%		
24) Chloroform-d	4.249	84	138831	44.231	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.460%		
26) 1,2-Dichloroethane-d4	4.940	65	95841	47.258	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.520%		
32) Benzene-d6	4.960	84	278338	45.176	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	90.360%		
36) 1,2-Dichloropropane-d6	5.989	67	86342	42.574	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	85.140%		
41) Toluene-d8	7.243	98	254852	45.909	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.820%		
43) trans-1,3-Dichloroprop...	7.551	79	43333	43.233	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	86.460%		
47) 2-Hexanone-d5	8.021	63	61839	73.510	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	73.510%		
56) 1,1,2,2-Tetrachloroeth...	10.152	84	106187	41.014	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	82.020%		
66) 1,2-Dichlorobenzene-d4	11.561	152	98672	46.846	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.700%		
Target Compounds						Qvalue
5) Vinyl chloride	1.285	62	15670	7.581	ug/L	88
20) cis-1,2-Dichloroethene	3.821	96	3141	1.843	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012424\
Data File : VV033846.D
Acq On : 24 Jan 2024 12:21
Operator : SY/MD
Sample : P1191-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMQ3

Quant Time: Jan 25 00:57:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 25 00:55:39 2024
Response via : Initial Calibration

