

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012424\
 Data File : VV033854.D
 Acq On : 24 Jan 2024 17:31
 Operator : SY/MD
 Sample : P1191-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jan 25 01:10:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 25 00:55:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	194941	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	180709	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	90418	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	68623	45.803	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.600%
7) Chloroethane-d5	1.529	69	55457	46.741	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.480%
11) 1,1-Dichloroethene-d2	2.060	65	30965	48.024	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.040%
21) 2-Butanone-d5	3.783	46	86104	83.788	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.790%
24) Chloroform-d	4.249	84	121842	45.481	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.960%
26) 1,2-Dichloroethane-d4	4.941	65	86847	50.173	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.340%
32) Benzene-d6	4.960	84	237390	46.208	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.420%
36) 1,2-Dichloropropane-d6	5.989	67	75330	44.546	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.100%
41) Toluene-d8	7.243	98	216911	46.861	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.720%
43) trans-1,3-Dichloroprop...	7.555	79	37896	45.343	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.680%
47) 2-Hexanone-d5	8.021	63	58664	83.632	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.630%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	99192	45.946	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.900%
66) 1,2-Dichlorobenzene-d4	11.561	152	83691	48.786	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012424\
Data File : VV033854.D
Acq On : 24 Jan 2024 17:31
Operator : SY/MD
Sample : P1191-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jan 25 01:10:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 25 00:55:39 2024
Response via : Initial Calibration

