

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012822\
 Data File : VV024478.D
 Acq On : 28 Jan 2022 10:51
 Operator : SY/MD
 Sample : VV0128WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK298

Quant Time: Jan 29 02:24:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 02:05:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	156468	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	167820	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	84694	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	55952	4.082	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	81.600%	
7) Chloroethane-d5	1.568	69	56425	4.719	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.400%	
11) 1,1-Dichloroethene-d2	2.108	63	90147	3.441	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.800%	
20) 2-Butanone-d5	3.911	46	78928	36.127	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	72.260%	
24) Chloroform-d	4.342	84	132454	5.092	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.800%	
26) 1,2-Dichloroethane-d4	5.030	65	63032	5.350	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.000%	
32) Benzene-d6	5.046	84	244640	4.881	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.600%	
36) 1,2-Dichloropropane-d6	6.066	67	78072	4.947	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.000%	
41) Toluene-d8	7.313	98	201610	4.384	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.600%	
43) trans-1,3-Dichloroprop...	7.622	79	26146	4.824	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	96.400%	
46) 2-Hexanone-d5	8.088	63	94361	50.920	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.840%	
56) 1,1,2,2-Tetrachloroeth...	10.213	84	48466	5.159	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	77509	5.337	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.800%	
Target Compounds					Qvalue	
25) Chloroform	4.371	83	5155	0.216	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012822\
Data File : VV024478.D
Acq On : 28 Jan 2022 10:51
Operator : SY/MD
Sample : VV0128WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK298

Quant Time: Jan 29 02:24:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 02:05:31 2022
Response via : Initial Calibration

