

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012822\
 Data File : VW024493.D
 Acq On : 28 Jan 2022 18:01
 Operator : SY/MD
 Sample : N1328-03 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COHM2

Quant Time: Jan 29 03:52:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 03:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	123189	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	135297	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63441	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	42458	3.934	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.600%	
7) Chloroethane-d5	1.568	69	42741	4.540	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.800%	
11) 1,1-Dichloroethene-d2	2.117	63	357119	17.312	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	346.200%#	
20) 2-Butanone-d5	3.911	46	73568	42.770	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	85.540%	
24) Chloroform-d	4.352	84	101436	4.953	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.000%	
26) 1,2-Dichloroethane-d4	5.037	65	48006	5.176	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.600%	
32) Benzene-d6	5.053	84	179706	4.447	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
36) 1,2-Dichloropropane-d6	6.072	67	61947	4.868	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.400%	
41) Toluene-d8	7.316	98	144708	3.903	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	78.000%	
43) trans-1,3-Dichloroprop...	7.628	79	18219	4.170	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.400%	
46) 2-Hexanone-d5	8.091	63	79675	53.330	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	106.660%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	38782	5.121	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.400%	
66) 1,2-Dichlorobenzene-d4	11.625	152	54430	5.003	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.000%	
Target Compounds						Qvalue
5) Vinyl chloride	1.310	62	49368	3.898	ug/L	98
12) 1,1-Dichloroethene	2.117	96	552966	59.732	ug/L #	78
19) 1,1-Dichloroethane	3.191	63	5922	0.358	ug/L	92
22) cis-1,2-Dichloroethene	3.911	96	194182	21.764	ug/L	93
25) Chloroform	4.384	83	5389	0.287	ug/L	91
29) 1,1,1-Trichloroethane	4.609	97	139156	8.867	ug/L	100
34) Trichloroethene	5.911	95	1898987	185.520	ug/L	93
45) 1,1,2-Trichloroethane	7.841	97	894	0.149	ug/L #	67
47) Tetrachloroethene	7.976	164	582756	72.272	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012822\
Data File : VV024493.D
Acq On : 28 Jan 2022 18:01
Operator : SY/MD
Sample : N1328-03 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0HM2

Quant Time: Jan 29 03:52:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 03:50:16 2022
Response via : Initial Calibration

