

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012822\
 Data File : VW024495.D
 Acq On : 28 Jan 2022 18:48
 Operator : SY/MD
 Sample : N1328-05 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COHM4

Quant Time: Jan 29 03:52:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 03:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	124349	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	132471	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62041	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	41304	3.791	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	75.800%	
7) Chloroethane-d5	1.568	69	42649	4.488	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.800%	
11) 1,1-Dichloroethene-d2	2.108	63	68723	3.300	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	66.000%	
20) 2-Butanone-d5	3.918	46	55807	32.142	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	64.280%	
24) Chloroform-d	4.349	84	100735	4.873	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.400%	
26) 1,2-Dichloroethane-d4	5.034	65	48917	5.225	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.400%	
32) Benzene-d6	5.050	84	177397	4.484	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.600%	
36) 1,2-Dichloropropane-d6	6.072	67	60210	4.833	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.600%	
41) Toluene-d8	7.317	98	139755	3.850	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.000%	
43) trans-1,3-Dichloroprop...	7.625	79	18158	4.244	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.800%	
46) 2-Hexanone-d5	8.091	63	72664	49.675	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.340%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39128	5.277	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.600%	
66) 1,2-Dichlorobenzene-d4	11.625	152	55026	5.172	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.400%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.114	96	2565	0.274	ug/L	# 1
19) 1,1-Dichloroethane	3.195	63	3226	0.193	ug/L	99
22) cis-1,2-Dichloroethene	3.912	96	12348	1.371	ug/L	96
25) Chloroform	4.381	83	4502	0.237	ug/L	95
34) Trichloroethene	5.921	95	11408	1.138	ug/L	93
47) Tetrachloroethene	7.976	164	93194	11.804	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012822\
Data File : VV024495.D
Acq On : 28 Jan 2022 18:48
Operator : SY/MD
Sample : N1328-05 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0HM4

Quant Time: Jan 29 03:52:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 03:50:16 2022
Response via : Initial Calibration

