

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012822\
 Data File : VW024497.D
 Acq On : 28 Jan 2022 19:35
 Operator : SY/MD
 Sample : N1328-07 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COHM6

Quant Time: Jan 29 03:53:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 03:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	109069	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	120546	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60816	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	40185	4.206	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	84.200%	
7) Chloroethane-d5	1.568	69	43698	5.243	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.800%	
11) 1,1-Dichloroethene-d2	2.108	63	71950	3.939	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	78.800%	
20) 2-Butanone-d5	3.911	46	58423	38.363	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	76.720%	
24) Chloroform-d	4.352	84	95964	5.292	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.800%	
26) 1,2-Dichloroethane-d4	5.034	65	47375	5.769	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.400%	
32) Benzene-d6	5.053	84	164587	4.572	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.400%	
36) 1,2-Dichloropropane-d6	6.072	67	58792	5.186	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.800%	
41) Toluene-d8	7.316	98	123261	3.732	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	74.600%	
43) trans-1,3-Dichloroprop...	7.628	79	15748	4.045	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.000%	
46) 2-Hexanone-d5	8.091	63	65538	49.235	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.480%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	36459	5.403	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	108.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	46019	4.413	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	88.200%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.117	96	6844	0.835	ug/L	# 1
16) Methylene chloride	2.513	84	1629	0.194	ug/L	83
25) Chloroform	4.374	83	4452	0.267	ug/L	98
29) 1,1,1-Trichloroethane	4.609	97	5649	0.404	ug/L	96
34) Trichloroethene	5.924	95	11154	1.223	ug/L	95
47) Tetrachloroethene	7.976	164	237007	32.990	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012822\
Data File : VV024497.D
Acq On : 28 Jan 2022 19:35
Operator : SY/MD
Sample : N1328-07 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COHM6

Quant Time: Jan 29 03:53:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 03:50:16 2022
Response via : Initial Calibration

