

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012822\
 Data File : VW024505.D
 Acq On : 28 Jan 2022 22:41
 Operator : SY/MD
 Sample : N1328-15 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0HN4

Quant Time: Jan 29 03:55:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 03:50:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	101711	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	111030	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	50623	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	39380	4.419	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.568	69	42541	5.473	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.400%
11) 1,1-Dichloroethene-d2	2.108	63	71755	4.213	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.200%
20) 2-Butanone-d5	3.908	46	70723	49.799	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.600%
24) Chloroform-d	4.349	84	92934	5.496	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.000%
26) 1,2-Dichloroethane-d4	5.037	65	45686	5.966	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.400%
32) Benzene-d6	5.053	84	153842	4.639	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	6.069	67	57106	5.469	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.400%
41) Toluene-d8	7.317	98	115798	3.806	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.200%
43) trans-1,3-Dichloroprop...	7.629	79	15567	4.341	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.800%
46) 2-Hexanone-d5	8.092	63	64265	52.417	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.840%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35577	5.724	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	43997	5.068	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%
Target Compounds						
12) 1,1-Dichloroethene	2.121	96	5726	0.749	ug/L #	1
16) Methylene chloride	2.507	84	1518	0.194	ug/L	91
19) 1,1-Dichloroethane	3.191	63	1059	0.077	ug/L #	60
22) cis-1,2-Dichloroethene	3.918	96	15590	2.116	ug/L	84
25) Chloroform	4.381	83	4492	0.289	ug/L	100
29) 1,1,1-Trichloroethane	4.613	97	4362	0.339	ug/L #	89
34) Trichloroethene	5.918	95	20059	2.388	ug/L	93
47) Tetrachloroethene	7.976	164	67987	10.274	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012822\
Data File : VV024505.D
Acq On : 28 Jan 2022 22:41
Operator : SY/MD
Sample : N1328-15 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0HN4

Quant Time: Jan 29 03:55:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 03:50:16 2022
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 0.749 ug/L

RT: 2.121 min Scan# 31

Delta R.T. 0.003 min

Lab File: VV024505.D

Acq: 28 Jan 2022 22:41

Instrument :

MSVOA_V

ClientSampleId :

COHN4





Tgt Ion: 96 Resp: 5726

Ion Ratio Lower Upper

96 100

61 242.3 116.1 215.5#

63 670.8 79.4 147.4#



#16

Methylene chloride

Concen: 0.194 ug/L

RT: 2.507 min Scan# 456

Delta R.T. -0.000 min

Lab File: VV024505.D

Acq: 28 Jan 2022 22:41




Tgt Ion: 84 Resp: 1518


Ion Ratio Lower Upper

84 100

86 53.3 42.6 79.0

49 122.7 79.7 147.9









#19
1,1-Dichloroethane
Concen: 0.077 ug/L
RT: 3.191 min Scan# 60
Delta R.T. 0.003 min
Lab File: VV024505.D
Acq: 28 Jan 2022 22:41




Instrument :
MSVOA_V
ClientSampleId :
COHN4


Tgt Ion: 63 Resp: 1059
Ion Ratio Lower Upper
63 100
65 55.2 22.3 41.3#
83 0.0 10.7 19.9#



#22
cis-1,2-Dichloroethene
Concen: 2.116 ug/L
RT: 3.918 min Scan# 895
Delta R.T. 0.010 min
Lab File: VV024505.D
Acq: 28 Jan 2022 22:41

Tgt Ion: 96 Resp: 15590
Ion Ratio Lower Upper
96 100
61 137.2 83.4 154.8
68 0.0 0.0 0.0





#25

Chloroform

Concen: 0.289 ug/L

RT: 4.381 min Scan# 1109

Delta R.T. 0.010 min

Lab File: VV024505.D

Acq: 28 Jan 2022 22:41

Instrument :

MSVOA_V

ClientSampleId :





COHN4

Tgt Ion: 83 Resp: 4492

Ion Ratio Lower Upper

83 100

85 65.0 45.2 84.0



#29

1,1,1-Trichloroethane

Concen: 0.339 ug/L

RT: 4.613 min Scan# 1111

Delta R.T. 0.010 min

Lab File: VV024505.D

Acq: 28 Jan 2022 22:41




Tgt Ion: 97 Resp: 4362


Ion Ratio Lower Upper

97 100

99 60.8 51.9 77.9

61 53.7 33.2 49.8#









#34
Trichloroethene
Concen: 2.388 ug/L
RT: 5.918 min Scan# 11
Delta R.T. 0.006 min
Lab File: VV024505.D
Acq: 28 Jan 2022 22:41

Instrument :
MSVOA_V
ClientSampleId :
COHN4

Tgt Ion: 95 Resp: 20059
Ion Ratio Lower Upper
95 100
97 64.3 45.2 84.0
132 97.1 76.1 141.3
130 101.1 75.4 140.0



#47
Tetrachloroethene
Concen: 10.274 ug/L
RT: 7.976 min Scan# 2157
Delta R.T. 0.003 min
Lab File: VV024505.D
Acq: 28 Jan 2022 22:41

Tgt Ion:164 Resp: 67987
Ion Ratio Lower Upper
164 100
129 94.7 62.6 116.2
131 89.4 61.0 113.4
166 130.5 87.3 162.1

